



Declaration of conformity

For the following equipment :

Product Name: AC/DC Medical Adaptor

Model Designation: GSM90Ax (x=12,15,19,24,48)

is herewith confirmed to comply with the requirements set out in the Council Directive, the following standards were applied :

RoHS Directive (2011/65/EU), (EU)2015/863

MDR Directive (EU) 2017/745

EN 60601-1:2006+A11+A1+A12

TUV certificate No : TA 50341433

MDR Directive (EU) 2017/745

EN 60601-1-2:2015

EMI (Electro-Magnetic Interference)

Conducted emission / Radiated emission

| | | |
|------------------|-------------------------------|---------|
| | EN 55011:2016+ A11:2020 | |
| | EN IEC 61024-3:2018 | Class B |
| Harmonic current | EN IEC 61000-3-2:2019+A1:2021 | |
| Voltage flicker | EN 61000-3-3:2013+A1:2019 | |

EMS (Electro-Magnetic Susceptibility)

EN 60601-1-2:2015 EN IEC 61024-3:2018

| | | | |
|--------------------------|---------------------------|---------|--------------------------|
| ESD air | EN 61000-4-2:2009 | Level 4 | 15KV |
| ESD contact | EN 61000-4-2:2009 | Level 4 | 8KV |
| RF field susceptibility | EN IEC 61000-4-3:2020 | Level 3 | 10V/m(80MHz~2.7GHz) |
| RF field susceptibility | EN IEC 61000-4-3:2020 | Table 9 | 9~28V/m (385MHz~5.78GHz) |
| EFT bursts | EN 61000-4-4:2012 | Level 3 | 2KV/100KHz |
| Surge susceptibility | EN 61000-4-5:2014+A1:2017 | Level 3 | 1KV/Line-Line |
| Surge susceptibility | EN 61000-4-5:2014+A1:2017 | Level 3 | 2KV/Line-FG |
| Conducted susceptibility | EN 61000-4-6:2014 | Level 3 | 10V |
| Magnetic field immunity | EN 61000-4-8:2010 | Level 4 | 30A/m |

Voltage dip, interruption EN IEC 61000-4-11:2020
0% residual voltage for 0.5 cycles, 0% residual voltage for 1 cycles, 40% residual voltage for 10 cycles, 70% residual voltage for 25 cycles, 80% residual voltage for 250 cycles, 0% residual voltage for 250 cycles

Note:

The power supply is considered as a component that will be operated in combination with final equipment. Since EMC performance will be affected by the complete system, the final equipment manufacturers must re-qualify EMC Directive on the complete system again.

For guidance on how to perform these EMC tests, please refer to TDF (Technical Documentation File).

The product under declaration is just a unit without medical function. Complete MDR should only be verified when it is used together with particular medical device(s).

This Declaration is effective from serial number SC1xxxxxxx

Person responsible for marking this declaration :

Mean Well Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 24891, Taiwan

(Manufacturer Address)

Aries Jian/ Director, Group R&D :

(Name / Position)

(Signature)

Alex Tsai/Director, Product Strategy Center :

(Name / Position)

(Signature)

Taiwan

(Place)

September 14, 2021

(Date)